



## Mouse Interleukin 3 ELISA Kit (2 plates)

**Catalogue No.:** BEK-2037-2P

**Description:** The mouse Interleukin 3 (IL-3) Kit is a sandwich ELISA. The capture antibody is a polyclonal mouse IL-3 antibody pre-coated onto the 96-well strip plates provided in the kit. Mouse test samples and standards of known IL-3 concentration are added to these wells and allowed to complex with the bound IL-3 antibody. A biotinylated mouse IL-3 polyclonal antibody is then added. This detection antibody binds to the antigen thus completing the sandwich. After washing, an enzyme Avidin-Biotin-Peroxidase complex (ABC) is added which binds to the second antibody. The peroxidase substrate TMB is added to induce a coloured reaction product. The intensity of this coloured product is directly proportional to the concentration of IL-3 present in the samples. The purpose of this kit is the in-vitro quantitative determination of mouse IL-3 in samples such as sera, plasma, tissue lysates and cell culture supernates. This kit has been configured for research use only and is not to be used in diagnostic or clinical procedures.

**Batch No.:** See product labels

**Other Names:** Interleukin-3; IL-3; IL3; Hematopoietic growth factor; Mast cell growth factor; MCGF; Multipotential colony-stimulating factor; P-cell-stimulating factor; IL3;Csfmu; IL-3;

**Accession:** P01586 IL3\_MOUSE;

**Specificity:** Mouse IL-3

**Storage:** Store at 4°C

**Kit components:** The ELISA kit box contains 2 x 96-well pre-coated strip plates, protein standards, detection reagents, substrate buffer and detailed protocols.


**Range:** 7.8 pg/ml - 500 pg/ml

**Sensitivity:** < 1 pg/ml

**Kit protocol:** Please refer to our online product listing for current protocol/MSDS versions.

Typical Mouse IL-3 ELISA Kit Standard Curve  
(TMB reaction incubated at 37°C for 14 min)

Concentration (pg/ml)	0.0	7.8	15.6	31.3	62.5	125	250	500
O.D.	0.060	0.150	0.217	0.370	0.736	1.405	2.468	2.954



This standard curve is for demonstration purposes only. A standard curve should be generated for each assay.

FOR RESEARCH USE ONLY