

## Mouse C-C motif chemokine 22 ELISA Kit (2 plates)

**Catalogue No.:** BEK-2063-2P

**Description:** The mouse C-C motif chemokine 22 (CCL22) or Macrophage-derived chemokine (MDC) Kit is a sandwich ELISA. The capture antibody is a monoclonal mouse MDC antibody pre-coated onto the 96-well strip plates provided in the kit. Mouse test samples and standards of known MDC concentration are added to these wells and allowed to complex with the bound MDC antibody. A biotinylated mouse MDC polyclonal antibody is then added. This detection antibody binds to the antigen thus completing the sandwich. After washing, an enzyme Avidin-Biotin-Peroxidase complex (ABC) is added which binds to the second antibody. The peroxidase substrate TMB is added to induce a coloured reaction product. The intensity of this coloured product is directly proportional to the concentration of MDC present in the samples. The purpose of this kit is the in-vitro quantitative determination of mouse MDC in samples such as sera, plasma, tissue lysates and cell culture supernates. This kit has been configured for research use only and is not to be used in diagnostic or clinical procedures.

**Batch No.:** See product labels

**Other Names:** Activated B and dendritic cell-derived; CC chemokine ABCD-1; Small-inducible cytokine A22; Ccl22; Abcd1; Scya22; MDC;

**Accession:** O88430 CCL22\_MOUSE;

**Specificity:** Mouse MDC

**Storage:** Store at 2-8C

**Kit components:** The ELISA kit box contains 2 x 96-well pre-coated strip plates, protein standards, detection reagents, substrate buffer and detailed protocols.


**Range:** 15.6 pg/mL - 1,000 pg/mL

**Sensitivity:** < 1 pg/ml

**Kit protocol:** Please refer to our online product listing for current protocol/MSDS versions.

**Typical Mouse MDC ELISA Kit Standard Curve**  
(TMB reaction incubated at 37°C for 12 min)

Concentration (pg/ml)	0.0	15.6	31.3	62.5	125	250	500	1000
O.D.	0.015	0.078	0.144	0.302	0.648	1.333	1.977	2.439



This standard curve is for demonstration purposes only. A standard curve should be generated for each assay.

FOR RESEARCH USE ONLY