

Human Resistin ELISA Kit (2 plates)

Catalogue No.: BEK-2115-2P

Description: The human Resistin Kit is a sandwich ELISA. The capture antibody is a polyclonal human Resistin antibody pre-coated onto the 96-well strip plates provided in the kit. Human test samples and standards of known Resistin concentration are added to these wells and allowed to complex with the bound Resistin antibody. A biotinylated human Resistin polyclonal antibody is then added. This detection antibody binds to the antigen thus completing the sandwich. After washing, an enzyme Avidin-Biotin-Peroxidase complex (ABC) is added which binds to the second antibody. The peroxidase substrate TMB is added to induce a coloured reaction product. The intensity of this coloured product is directly proportional to the concentration of Resistin present in the samples. The purpose of this kit is the in-vitro quantitative determination of human Resistin in samples such as sera, plasma, tissue lysates and cell culture supernates. This kit has been configured for research use only and is not to be used in diagnostic or clinical procedures.

Batch No.: See product labels

Other Names: Adipose tissue-specific secretory factor; ADSF; C/EBP-epsilon-regulated myeloid-specific secreted cysteine-rich protein; Cysteine-rich secreted protein A12-alpha-like 2; Cysteine-rich secreted protein FIZZ3; RETN; FIZZ3; HXCP1; RSTN;

Accession: Q9HD89 RETN_HUMAN;

Specificity: Human Resistin

Storage: Store at 4°C

Kit components: The ELISA kit box contains 2 x 96-well pre-coated strip plates, protein standards, detection reagents, substrate buffer and detailed protocols.


Range: 78 pg/ml - 5,000 pg/ml

Sensitivity: < 3 pg/ml

Kit protocol: Please refer to our online product listing for current protocol/MSDS versions.

Typical Human Resistin ELISA Kit Standard Curve
(TMB reaction incubated at 37°C for 30 min)

Concentration (pg/ml)	0.0	156	313	625	1250	2500	5000	10,000
O.D.	0.013	0.317	0.574	0.983	1.509	1.988	2.045	2.139



This standard curve is for demonstration purposes only. A standard curve should be generated for each assay.

FOR RESEARCH USE ONLY