

**biosensis<sup>®</sup> Mouse Interleukin-4 (IL-4)/  
B-cell stimulatory factor 1 (BSF-1)/ Lymphocyte  
stimulatory factor 1 ELISA  
Kit Protocol**

**Catalog Number: BEK-2039-2P**

*For quantitative detection of mouse IL-4 in cell culture supernatant and serum samples only  
when used as directed, other substrates not tested.*

**TABLE OF CONTENTS**

|             |  |   |
|-------------|--|---|
| <b>I</b>    | Materials provided.....                  | 2 |
| <b>II</b>   | Equipment required but not supplied..... | 2 |
| <b>III</b>  | Technical hints .....                    | 2 |
| <b>IV</b>   | Storage of kit components.....           | 3 |
| <b>V</b>    | Sample preparation and dilution.....     | 4 |
| <b>VI</b>   | Reagent preparation.....                 | 5 |
| <b>VII</b>  | Assay procedure.....                     | 6 |
| <b>VIII</b> | Calculation of results.....              | 7 |
| <b>IX</b>   | Typical Standard curve.....              | 8 |

---

*For Research Use Only. Not for use in Clinical & Diagnostic Procedures*

1

---

## I. Materials Provided

| <b>Kit Components</b>   | <b>Unit Size/Volume</b> |
|---|-------------------------|
| Lyophilized recombinant mouse IL-4 standard   | 10ng/tube x 4           |
| 96 wells plate pre-coated with anti-mouse IL-4 antibody (12 strips of 8 wells =1 plate) | 2 plates                |
| Sample diluent buffer   | 30ml x 2                |
| Biotinylated anti- mouse IL-4 antibody (dilution 1:100)                                 | 100µl x 2               |
| Antibody diluent buffer   | 12ml x 2                |
| Avidin-Biotin-Peroxidase Complex (ABC) (dilution 1:100)                                 | 100µl x 2               |
| ABC diluent buffer  | 12ml x 2                |
| TMB colour developing agent   | 10ml x 2                |
| TMB stop solution   | 10ml x 2                |
| Wash Buffer Concentrate (25x)   | 20ml x 2                |
| Plate Sealers   | 1 ea x 8                |

**WARNING: DO NOT MIX REAGENTS SUCH AS ANTIBODY BUFFERS ETC BETWEEN DIFFERENT ASSAYS IF USING MULTIPLE KITS. USE ONLY THE BUFFERS, STANDARDS & SOLUTIONS INCLUDED FOR YOUR SPECIFIC LOT. MIXING SOLUTIONS FROM DIFFERENT TARGET KITS CAN LEAD TO POOR KIT PERFORMANCE**

---

## II. Equipment Required but Not Supplied

1. Distilled Water, 1000 mL +, graduated cylinders for measuring
  2. Automated plate washer is desirable
  3. Multi-channel or repeating pipette.
  4. Clean tubes and Eppendorf tubes.
  5. Plate shaker (300-500 rpm)
  6. Mechanical Vortex.
  7. Microplate reader with 450nm filter.
- 

## III. Technical Hints

1. Spin all kit components to get liquid to the bottom of tubes before use.
2. Duplicate well assays are recommended for both standard and sample testing.

3. In order to avoid marginal effects of plate incubation due to temperature differences, it is suggested that the ABC and TMB solutions be brought to room temperature (or 37° C if using option protocol) 30 min before use.
4. Recommended method for manual plate washing.
  - Emptying of wells on the plate:  
Place the plate on the palm of the hand in a position that enables easy flicking movement using the wrist. Holding the plate over a sink, quickly invert the plate, whilst accelerating the arm downward toward the sink. Abruptly stop the downward acceleration to force the liquid from the wells into the sink. When done correctly the technique should prevent liquid from getting on to the fingers or on the outside of the strip wells or plate holder. *Note: Retain the upside down position of the plate to avoid any back flow into the wells. DO NOT LET THE PLATE DRY OUT.*
  - Blotting the plate:  
Immediately blot the inverted plate by lightly tapping the plate 3-4 times on blotting paper.
  - Washing: Forcefully pipette Wash Buffer into each well with a multi-channel pipette. Empty the wells of wash buffer using technique described above. Repeat washing and flicking procedures thrice. Wash plate with 1X Wash buffer, and each time let washing buffer stay in the wells for 1-2 min before rinsing with more wash buffer as directed above. *Note: Avoid touching the inside surface of the wells with the pipette tips.*

***Do not let the wells dry out at any time or enzymatic activity will be lost.***

5. Preparation of Wash buffer:
  - Prepare 500 ml of Working Wash Buffer by diluting the supplied 20 ml of Wash Buffer (25 x) with 480 ml of deionized or distilled water. If crystals have formed in the concentrate, warm to room temperature and mix it gently until crystals have completely dissolved.  
*Note: if blank values are over 0.2 the addition of 0.05% Tween-20 PBS wash buffers will usually help lower them as well as overall assay backgrounds and can help improve signal/noise ratios.*

#### **IV. Storage of Kit Components**

This kit may be stored for up to 6 months at 2°C to 8°C from the date of shipment. Standards should be stored at -20°C or -80°C (recommended at -80°C) after reconstitution. Coated strips or reagents may be stored for up to 1 month at 2°C to 8°C. Return unused wells to the pack and reseal the pack. *Note: The kit can be used within 8 months if the whole kit is stored at -20°C. Avoid repeated freeze-thaw cycles.*

## V. Sample Preparation and Dilution

- **Cell culture supernatant:** Remove particulates by centrifugation (10,000 x g 5 min), analyze immediately or aliquot and store at -20°C-80°C. Divide samples into aliquots to prevent multiple freeze thaws; use appropriate protease inhibitor cocktails for best results.
- **Serum:** Allow the serum to clot in a serum separator tube (about 4 hours) at room temperature. Centrifuge at approximately 1000 X g for 15 min. Analyze the serum immediately or aliquot and store frozen at -20°C-80°C. **Not suitable for serum which is haemolytic or with high fat.**

*Note: The user needs to estimate the concentration of the target protein in the sample and select a proper dilution factor so that the diluted target protein concentration falls near the middle of the linear regime in the standard curve. Also with certain growth factors and target proteins, binding proteins and location can influence over all signal. It is recommended that researchers confirm by literature an effective lysis, cell membrane preparations and/or extraction buffer formulations and treatments that are suitable for the precise target being assayed.*

The following table can be used as a guideline for sample dilution. Dilute the sample using the provided diluent buffer. **Mix the sample thoroughly with diluent buffer.**

| target protein concentration                 | Concentration range  | Working dilution      | Amount of sample diluent buffer to be added       |
|--|----------------------|-----------------------|---|
| <b>High target protein concentration</b>     | <b>5-50 ng/ml</b>    | 1:100                 | Add 1 µl sample into 99 µl sample diluent buffer  |
| <b>Medium target protein concentration</b>   | <b>0.5-5 ng/ml</b>   | 1:10                  | Add 10 µl sample into 90 µl sample diluent buffer |
| <b>Low target protein concentration</b>      | <b>7.8-500 pg/ml</b> | 1:2                   | Add 50 µl sample to 50 µl sample diluent buffer   |
| <b>Very Low target protein concentration</b> | <b>≤ 7.8 pg/ml</b>   | No dilution necessary | (working dilution is 1:2)                         |

Range: 7.8 pg/ml to 500 pg/mL

Sensitivity: <1.0 pg/mL

Specificity: Natural and recombinant mouse IL-4

Cross-reactivity: No detectable cross reactivity with other relevant proteins. Rat IL-4 protein sequence shares only 61% homology but is untested in this assay.

| Sample       | Intra-Assay Precision |      |       | Inter-Assay Precision |      |      |
|--------------|-----------------------|------|-------|-----------------------|------|------|
|              | 1                     | 2    | 3     | 1                     | 2    | 3    |
| N            | 16                    | 16   | 16    | 24                    | 24   | 24   |
| Mean (pg/ml) | 15                    | 96   | 280   | 16                    | 93   | 275  |
| Std. Dev.    | 1.11                  | 6.62 | 20.44 | 1.48                  | 6.69 | 20.9 |
| CV(%)        | 7.4                   | 6.9  | 7.3   | 9.3                   | 7.2  | 7.6  |

## I. Reagent Preparation

### 1. Preparation of Standards

- Reconstitution of standard with the sample diluent provided.
  - (a) 10,000 pg/ml (10 ng/ml) of mouse IL-4 standard solution: Add 1ml of sample diluent buffer into **one** tube of standard, mix thoroughly and keep the tube at room temperature for 10 min. Note: The standard product used in this kit is recombinant mouse IL-4 protein comprising of amino acids 21-140 expressed in and purified from an *E.coli*. expression system.
  - (b) 500 pg/ml of mouse IL-4 standard solution: Add 0.050 ml (50µL) of the above 10,000 pg/ml IL-4 standard solution into 0.95 ml sample diluent buffer and mix thoroughly.
  - (c) 250 pg/ml to 7.8 pg/ml of mouse IL-4 standard solutions: Label 6 Eppendorf tubes with 250 pg/ml, 125 pg/ml, 62.5 pg/ml, 31.25 pg/ml, 15.625 pg/ml, 7.8 pg/ml respectively. Aliquot 0.3 ml of sample diluent buffer into each tube. THEN Add 0.3 ml of the above 500 pg/ml standard solution (item a) into the 1st tube and mix. Transfer 0.3ml from 1<sup>st</sup> tube into 2<sup>nd</sup> and mix. Continue performing serial dilutions.
- *Note: Standard solution should be prepared no more than 2 hours prior to the experiment. The working standard solution may be stored at 4°C for up to 12 hours, or at -20°C for up to 48 hours. Avoid repeated freeze-thaw cycles.*

### 2. Preparation of Biotinylated antibody working solution

- The total volume should be: 0.1ml/well x (the number of wells). (Allowing 0.1-0.2 ml more than total volume).
- Biotinylated antibody (Reagent B) should be diluted in 1:99 with the antibody diluent buffer and mixed thoroughly.
- *Note: the solution should be prepared no more than two hours prior to the experiment.*

### 3. Avidin-Biotin-Peroxidase Complex (ABC)

- The total volume should be: 0.1ml/well x (the number of wells). (Allowing 0.1-0.2 ml more than total volume).
- ABC should be diluted in 1:99 with the ABC diluent buffer and mixed thoroughly.
- *Note: the solution should be prepared no more than one hour prior to the experiment.*

---

## VII. Assay Procedure

We recommend that standard solution dilutions and each sample are plated in duplicate. Read entire protocol before beginning; if using optional method the ABC working solution and TMB color developing agent must be kept warm at 37°C for 30 min before use.

1. Aliquot out the standards into each well of the pre-coated plate. Add 0.1ml of the sample diluent buffer into the control well (Zero well). Add 0.1ml of each properly diluted sample into each empty well. We recommend at least duplicate standards and samples. *Note: See “VII. Sample Preparation and Dilution Guideline” above for details.*
2. Seal the plate with parafilm and incubate the plate at 4°C overnight. *Note: we recommend users leave the plate on a shaker if possible.* (Optional method: 37°C, 90 minutes but it is slightly less sensitive)
3. Remove the cover, discard the plate contents and blot the plate onto paper towels or other absorbent material. **DO NOT** let the wells completely dry at any time. Wash plate 3-5 times using wash buffer as described in the technical hints. *Note: See “V. Technical Hints - Recommended method for manual plate washing” above for details.*
4. Add 0.1ml of biotinylated antibody working solution into each well. Cover and incubate the plate at room temperature on a plate shaker for 2-3 hours. (Optional 37°C, 60 minutes with shaking) Wash 3X as described in the technical hints.
5. Add 0.1ml of prepared ABC working solution (*Note: See “VIII. Reagent Preparation” above for details*) into each well. Cover and incubate the plate at room temperature on a plate shaker for 1 hour. (Optional 37°C, 30 minutes with shaking). Wash 5X as described in technical hints allowing the wash buffer to remain in the wells for 1-2 minutes.
6. Warm TMB to room temperature (37° if using optional protocol). Add 90 ul of prepared TMB color developing agent into each well and incubate **in the dark** for 30-35 minutes at RT (or 15-30 minutes at 37°C). The optimal incubation time must be empirically determined. A guideline to look for is blue shading for the top four standard wells, while the remaining standards remain clear. Stopping the reaction too quickly can lead to low standard curve values. *Note: depending upon the sample dilution the experimental samples may show color much faster, if this is the case repeating the assay at a lower sample dilution is recommended if sample O.Ds are out of range.*
7. Stop the reaction by adding 0.1ml of prepared TMB stop solution into each well. The color will change into yellow.
8. Immediately read the plate at 450nm with a microplate reader. *Note: Color will fade over time; hence, we recommend plate to be read within 30 min after adding the stop solution.*

**CAUTION:** Bubbles in the wells will cause inaccurate readings. Ensure that all bubbles are removed prior to taking the absorbance reading.

*For Research Use Only. Not for use in Clinical & Diagnostic Procedures*

6

---

## VIII. Calculation of Results

### (a) Manual Plate Reading:

(The relative O.D.<sub>450</sub>) = (the O.D.<sub>450</sub> of each well) – (the O.D.<sub>450</sub> of Zero well).

The standard curve can be plotted as the relative O.D.<sub>450</sub> of each standard solution (Y) vs. the respective concentration of the standard solution (X). Known concentrations of the target protein are plotted on the X-axis and the corresponding O.D.<sub>450</sub> on the Y-axis. The standard curve should result in a graph that shows a direct relationship between target protein concentrations and the corresponding O.D.<sub>450</sub>. The greater the concentration of target protein in the sample, the higher the O.D.<sub>450</sub>.

*Note: if the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution.*

### Determine concentration of target protein in unknown sample:

The target protein concentration of the samples can be interpolated from the standard curve. Draw a horizontal line to intersect with the standard curve. A vertical line dropped from this point intersects the X-axis at the concentration of antibody in the unknown sample.

### (b) PC Interface Plate Reading

Enter the data into computer program curve fitting software. Use appropriate software to reduce the data and generate a four parameter logistic (4-PL) curve-fit; avoid using linear regression analysis. Perform a 4-PL regression analysis to calculate the concentration of target analyte in the samples. Multiply the result by the sample dilution factor. See our ELISA data analysis Technical Note on our website, [www.biosensis.com](http://www.biosensis.com) for more instructions on 4\_PL plotting.

Biosensis.com is your antibody and ELISA kit specialist  
Helping you make sense of your research!

We offer antibodies for research in Apoptosis, Autophagy, Cancer, Development, Diabetes & Obesity, DNA Damage, GPCRs, Growth Factors, Immunology, Matrix Molecules, Neuroscience and Stem Cells.

We have over 150 highly sensitive ELISA kits with a strip format offering the highest level of sensitivity and you only need to use what you need.

**Biosensis reagents are available online directly from [www.biosensis.com](http://www.biosensis.com)**



---

*For Research Use Only. Not for use in Clinical & Diagnostic Procedures*

7

### IX: Typical Standard Curve (for reference only, not to be used for actual data)

| Concentration<br>pg/ml | 0.0 pg/ml | 7.8 pg/ml | 15.6 pg/ml | 31.3 pg/ml | 62.5 pg/ml | 125 pg/ml | 250 pg/ml | 500 pg/ml |
|------------------------|-----------|-----------|------------|------------|------------|-----------|-----------|-----------|
| O.D.                   | 0.056     | 0.097     | 0.131      | 0.212      | 0.374      | 0.813     | 1.489     | 2.084     |



TMB reaction incubated at 37°C for 25-30 minutes